

No. D/21013/3284/19/08.07.2019/PW

GOVERNMENT OF INDIA
MINISTRY OF HOME AFFAIRS
PM DIVISION (PROCUREMENT WING)
JAISALMER HOUSE, 26-MAN SINGH ROAD,
NEW DELHI - 110 011
TEL NO. 91-011-23387306
Email: dc1.Procur-mha@gov.in

Dated, the 23rd December' 2019

OFFICE MEMORANDUM

Sub: Global Tender Enquiry No. D/21013/3284/19/08.07.2019/PW Dated 31.07.2019 Floated on CPP Portal and Opened on 06.09.2019 for Supply, Installation and Commissioning of "Brian Finger Printing / Brain Electrical Oscillation Signature Technology" Qty.01 unit for Central Forensic Science Laboratory CBI, New Delhi.

A GTE dated 31.07.2019 was floated on CPP Portal and was opened on 06.09.2019 (Tech Bid). The two firms viz. i) M/s Velankani Information Systems Ltd. and ii) M/s Axxonet System Technologies Pvt. Ltd. participated against the tender enquiry.

2. Tender Evaluation Committee (TEC) under the Chairmanship of Director, CFSL CBI, carried out technical evaluation of the bids submitted by the above two firms against the equipment mentioned in the tender enquiry. The TEC thoroughly examined the Qualification Criteria, eligibility criteria and evaluated the techno-commercial parameters of the bids documents submitted by both the above mentioned firms for their quoted items against the tender enquiry as per the Tender conditions. During evaluation the TEC unanimously decided to go for the field trial/ demonstration of both the firms for their quoted items from 06.11.2019 to 07.11.2019 at CFSL CBI, Block No. 4, CGO Complex, Lodhi Road, New Delhi-110003.

3. The TEC examined the respective equipment offered by both the participating firms against the tender enquiry to check the parameters as mentioned in the technical specifications of the GTE and the Corrigendum issued (if any). The following observations were made by the technical experts during the field trial:


23/12/19

NAME OF FIRM	REF. CLAUSE AS PER GTE	OBSERVATIONS
1. M/s Axxonet System Technologies Pvt. Ltd.	Appendix 2 clause 2 sub clause (a) and (c) at page 34 of GTE. a. Capacity to differentiate the experiential knowledge of the perpetrator, victim, accomplice and the witness related to the crime under reference c. Video and audio stimulus presentation facility	• The firm failed to differentiate the perpetrator and the witness using their system and software. In respect of first subject (witness), out of 90 probes presented, Experiential Knowledge was elicited in 21 probes in which 6 were false positive. In respect of the second subject (perpetrator), out of 90 probes presented, Experiential Knowledge was elicited in only 6 probes. The product is not having the video/image stimulus presentation facility thus, the product BEOS is not found suitable as per the requirements/technical specifications mentioned in the GTE.
2. M/s Velankani Information Systems Ltd.	Appendix 2 clause 2 sub clause (c), (d) and (e) at page 34 of GTE. c. Video and audio stimulus presentation facility d. Analysis software with the functions Viz. artefact rejection for all channels, Objective	• The software offered by the firm does not have the audio stimulus presentation facility. In addition, the reports generated are not system generated, but are manually prepared by the expert at the time of field trial. Hence, the offered equipment is not found suitable as per the requirements/

Handwritten signature
23/1/24

	analysis of the forensic data generated.	technical specifications mentioned in the GTE.
	e. Automatic analysis of the data acquired.	

4. Based on the examination of the bid documents submitted by both the participating firms and on the basis of the observations as analyzed by the TEC in the Field Trial, TEC recommended the following in its report: -

S.NO.	Name of Firm	Technically Qualified / Not Qualified
1.	M/s Axxonet System Technologies Pvt. Ltd.	NOT QUALIFIED
2.	M/s Velankani Information Systems Ltd	NOT QUALIFIED

5. As both the participating firms are "Technically Not Qualified" in the Technical Evaluation Report of TEC and since, no other participant left, hence, the current Global Tender Enquiry No. D/21013/3284/19/08.07.2019/PW dated 31.07.2019 reg. "Supply Installation and Commissioning of Brain Finger Printing/ Brain Electrical Oscillation Signature Technology" Qty. 01 unit for CFSL CBI is hereby **Scrapped**.

6. This has the approval of AS (PM).


23/7/19

(Inderjeet Chawla)
2IC (Proc)
PM Division, MHA

For and on behalf of President of India.

Copy to:

- i. The Director, CFSL CBI, Block No.4, CGO Complex, New Delhi-110003.
- ii. The Director, Women Safety Division/MHA, North Block, New Delhi.
- iii. The Section Officer (I.T.), IT cell website room no. 10, MHA, North Block, New Delhi, with the request to get it incorporated on the official website of MHA.
- iv. CPP Portal.
- v. M/s Velankani Information Systems Ltd., 43, Velankani Tech Park, Electronic City- Phase I, Hosur Road, Bengaluru-560100.
- vi. M/s Axxonet System Technologies Pvt. Ltd., 85/13, 4th Cross, Christ School Road, Bhavani Layout, Bangalore-560029.

Handwritten signature
23/11/20